
Midmark M11 - v1.0.5
Total Cycles: 4409

Name: RITTER2
Sterilizer ID: V2313228
03/06/2025 06:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 107.9 F 0.0
2:00 108.9 F 0.9
4:00 148.8 F 2.8
6:00 154.2 F 6.7
8:00 179.9 F 11.0
10:00 205.0 F 3.8
12:00 218.5 F 8.2
14:00 234.4 F 12.8
16:00 232.4 F 9.4
18:00 238.6 F 10.1
20:00 244.3 F 13.5
22:00 256.0 F 19.8
24:00 267.0 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.7
1:00 272.8 F 30.2
2:00 271.8 F 30.0
3:00 272.6 F 30.1
4:00 271.7 F 29.9

Min 271.6 F 29.6
Max 272.9 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 211.7 F 0.0
10:00 222.8 F 0.0
15:00 226.2 F 0.0
20:00 228.1 F 0.0
25:00 230.1 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:30
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:04:11
CYCLE COMPLETE

03/06/2025 07:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____