
Midmark M11 - v1.0.5
Total Cycles: 3561

Name: RITTER2
Sterilizer ID: V2313228
04/25/2024 03:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.4 F 0.0
2:00 94.7 F 0.5
4:00 102.4 F 2.2
6:00 122.9 F 4.8
8:00 155.6 F 8.1
10:00 177.7 F 11.7
12:00 209.4 F 4.5
14:00 221.1 F 8.6
16:00 228.3 F 12.9
18:00 232.3 F 9.0
20:00 240.1 F 11.5
22:00 242.1 F 12.3
24:00 253.6 F 18.0
26:00 263.9 F 24.2
28:00 272.1 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 30.0
1:00 272.0 F 30.0
2:00 271.9 F 30.0
3:00 272.1 F 29.9
4:00 272.1 F 30.0

Min 271.8 F 29.9
Max 272.8 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.0 F 0.0
5:00 224.6 F 0.0
10:00 228.5 F 0.0
15:00 236.2 F 0.0
20:00 219.6 F 0.0
25:00 216.6 F 0.0
30:00 216.9 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 28:54
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:07:26
CYCLE COMPLETE

04/25/2024 04:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____