
Midmark M11 - v1.0.5
Total Cycles: 3508

Name: RITTER2
Sterilizer ID: V2313228
04/09/2024 10:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.8 F 0.0
2:00 142.9 F 0.2
4:00 145.6 F 1.8
6:00 183.1 F 5.9
8:00 206.0 F 11.3
10:00 218.2 F 6.0
12:00 231.2 F 11.3
14:00 231.9 F 8.3
16:00 241.5 F 11.4
18:00 244.7 F 13.1
20:00 256.7 F 19.6
22:00 267.7 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.9 F 29.9
2:00 271.7 F 29.2
3:00 271.7 F 29.6
4:00 272.7 F 29.7

Min 271.1 F 29.0
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 209.4 F 0.0
10:00 207.3 F 0.0
15:00 197.3 F 0.0
20:00 193.4 F 0.0
25:00 192.4 F 0.0
30:00 191.4 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 23:24
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:01:45
CYCLE COMPLETE

04/09/2024 11:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____