
Midmark M11 - v1.0.0
Total Cycles: 3175

Name: RITTER 1
Sterilizer ID: V2178198
04/28/2023 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.9 F 0.0
2:00 196.2 F 3.1
4:00 205.8 F 8.2
6:00 224.1 F 9.9
8:00 232.5 F 9.9
10:00 237.4 F 8.2
12:00 245.1 F 12.8
14:00 245.2 F 12.7
16:00 258.2 F 19.7
18:00 269.7 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.1 F 28.7
2:00 271.4 F 28.4
3:00 272.1 F 28.7
4:00 271.3 F 28.1

Min 271.2 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 219.4 F 0.0
10:00 242.6 F 0.0
15:00 239.2 F 0.0
20:00 225.1 F 0.0
25:00 214.0 F 0.0
30:00 210.2 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 18:56
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:06
CYCLE COMPLETE

04/28/2023 10:56 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____