
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 008107
17:13 27/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:51
STERILIZING 4:00
136.4 C 223kPa 4:00
136.4 C 225kPa 4:30
136.4 C 222kPa 5:00
136.4 C 224kPa 5:30
136.1 C 223kPa 6:00
136.4 C 224kPa 6:30
136.0 C 221kPa 7:00
136.2 C 223kPa 7:30
136.2 C 223kPa 7:30
VENTING 7:31
AIR DRYING 8:24
CYCLE COMPLETE 9:03

Digital Signature #
A9726A112943BD7C
