
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000847
10:54 07/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:23
STERILIZING 4:38
136.0 C 220kPa 4:38
136.4 C 227kPa 5:08
136.4 C 232kPa 5:38
136.1 C 231kPa 6:08
135.9 C 231kPa 6:38
136.2 C 231kPa 7:08
136.1 C 231kPa 7:38
136.0 C 230kPa 8:08
136.0 C 230kPa 8:08
VENTING 8:08
AIR DRYING 9:00
CYCLE COMPLETE 24:00

Digital Signature #
616BD2729462EAB3
