
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 008500
15:15 02/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:25
STERILIZING 3:25
135.4 C 220kPa 3:25
136.1 C 222kPa 3:55
136.4 C 221kPa 4:25
136.4 C 223kPa 4:55
136.4 C 221kPa 5:25
136.3 C 224kPa 5:55
136.4 C 222kPa 6:25
135.9 C 220kPa 6:55
135.9 C 220kPa 6:55
VENTING 6:55
AIR DRYING 7:49
CYCLE COMPLETE 14:32

Digital Signature #
DDABF9AA7B750B84
