
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 008138
13:48 03/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:21
PRESSURIZING 3:22
STERILIZING 4:23
136.5 C 224kPa 4:23
136.3 C 224kPa 4:53
136.2 C 222kPa 5:23
136.4 C 224kPa 5:53
136.4 C 224kPa 6:23
136.4 C 222kPa 6:53
136.2 C 219kPa 7:23
136.3 C 221kPa 7:53
136.3 C 221kPa 7:53
VENTING 7:53
AIR DRYING 8:47
CYCLE COMPLETE 13:21

Digital Signature #
8E7A0AD21102D076
