
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000724
8:54 22/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:08
PRESSURIZING 2:16
STERILIZING 3:25
136.0 C 228kPa 3:25
136.4 C 233kPa 3:55
136.3 C 232kPa 4:25
136.2 C 230kPa 4:55
136.1 C 231kPa 5:25
136.2 C 231kPa 5:55
136.3 C 233kPa 6:25
136.4 C 233kPa 6:55
136.4 C 233kPa 6:55
VENTING 6:55
AIR DRYING 7:45
CYCLE COMPLETE 7:48

Digital Signature #
FAB20FBE61095AA9
