
Midmark M11 - v1.0.0
Total Cycles: 4766

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2024 01:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.7 F 0.0
2:00 154.1 F 1.6
4:00 151.7 F 5.2
6:00 179.9 F 9.4
8:00 208.2 F 8.9
10:00 222.1 F 6.0
12:00 233.8 F 10.6
14:00 229.3 F 6.7
16:00 241.8 F 11.7
18:00 238.1 F 9.4
20:00 251.9 F 16.0
22:00 263.6 F 22.8
24:00 271.5 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.4 F 28.5
2:00 272.0 F 28.6
3:00 272.2 F 28.4
4:00 271.9 F 28.6

Min 271.5 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 200.9 F 0.0
10:00 207.2 F 0.0
15:00 217.7 F 0.0
20:00 214.2 F 0.0
25:00 219.9 F 0.0
30:00 213.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:03
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:12
CYCLE COMPLETE

07/25/2024 02:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____