
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002500
12:04 10/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:41
STERILIZING 3:42
136.3 C 224kPa 3:42
136.3 C 225kPa 4:12
136.4 C 226kPa 4:42
136.4 C 226kPa 5:12
136.2 C 225kPa 5:42
136.2 C 226kPa 6:12
136.1 C 223kPa 6:42
136.3 C 226kPa 7:12
136.3 C 226kPa 7:12
VENTING 7:13
AIR DRYING 8:07
CYCLE COMPLETE 8:50

Digital Signature #
DCBD6F7B6F5E01F5
