
Midmark M11 - v1.0.0
Total Cycles: 3453

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2023 12:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.5 F 0.0
2:00 131.5 F 0.6
4:00 157.9 F 2.7
6:00 181.9 F 6.1
8:00 202.9 F 10.0
10:00 212.8 F 5.7
12:00 224.9 F 8.0
14:00 235.4 F 12.4
16:00 234.6 F 8.9
18:00 241.0 F 11.0
20:00 244.8 F 12.8
22:00 256.4 F 18.8
24:00 266.7 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 272.2 F 29.0
2:00 272.3 F 28.8
3:00 271.5 F 28.4
4:00 272.7 F 29.2

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.4 F 0.0
5:00 203.4 F 0.0
10:00 209.9 F 0.0
15:00 214.1 F 0.0
20:00 214.2 F 0.0
25:00 213.5 F 0.0
30:00 213.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 25:30
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:03:53
CYCLE COMPLETE

06/27/2023 01:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____