
Midmark M11 - v1.0.5
Total Cycles: 557

Name: RITTER2
Sterilizer ID: V2313228
11/17/2021 11:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.6 F 0.0
2:00 191.0 F 2.8
4:00 191.3 F 7.3
6:00 206.9 F 12.2
8:00 227.0 F 7.1
10:00 230.3 F 12.2
12:00 236.9 F 9.7
14:00 237.8 F 9.3
16:00 248.4 F 14.7
18:00 259.1 F 20.9
20:00 269.7 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.4
1:00 272.3 F 29.4
2:00 272.7 F 29.7
3:00 272.3 F 30.0
4:00 271.7 F 29.3

Min 271.5 F 29.3
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.2 F 0.0
5:00 231.9 F 0.0
10:00 226.8 F 0.0
15:00 222.9 F 0.0
20:00 201.3 F 0.0
25:00 188.8 F 0.0
30:00 188.1 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 21:50
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:00:59
CYCLE COMPLETE

11/17/2021 12:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____