
Midmark M11 - v1.0.0
Total Cycles: 1261

Name: RITTER 1
Sterilizer ID: V2178198
10/13/2021 11:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.8 F 0.0
2:00 168.6 F 2.3
4:00 178.6 F 5.8
6:00 207.0 F 10.8
8:00 222.5 F 5.1
10:00 234.7 F 10.4
12:00 237.3 F 7.8
14:00 244.1 F 13.0
16:00 242.4 F 11.5
18:00 255.3 F 18.0
20:00 266.6 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 29.0
2:00 271.5 F 28.2
3:00 272.3 F 29.1
4:00 272.2 F 28.7

Min 271.3 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 227.4 F 0.0
10:00 229.0 F 0.0
15:00 220.0 F 0.0
20:00 217.7 F 0.0
25:00 216.6 F 0.0
30:00 215.5 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:08
CYCLE COMPLETE

10/13/2021 12:01 PM
ACCEPTED: X
REJECTED:
Operator: U01