
Midmark M11 - v1.0.0
Total Cycles: 4295

Name: RITTER 1
Sterilizer ID: V2178198
02/12/2024 03:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.3 F 0.0
2:00 151.2 F 2.9
4:00 172.9 F 6.9
6:00 197.6 F 11.9
8:00 220.6 F 5.4
10:00 233.3 F 10.4
12:00 231.3 F 7.4
14:00 244.6 F 13.1
16:00 243.0 F 11.7
18:00 256.1 F 18.5
20:00 267.9 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.4 F 28.6
2:00 272.3 F 28.9
3:00 271.9 F 28.3
4:00 272.6 F 29.0

Min 271.5 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.0 F 0.0
5:00 211.8 F 0.0
10:00 220.3 F 0.0
15:00 223.7 F 0.0
20:00 217.6 F 0.0
25:00 216.8 F 0.0
30:00 213.0 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:14
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

02/12/2024 04:40 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____