
Midmark M11 - v1.0.5
Total Cycles: 4362

Name: RITTER2
Sterilizer ID: V2313228
02/20/2025 10:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 70.5 F 0.2
4:00 75.9 F 1.2
6:00 108.7 F 3.2
8:00 143.1 F 6.1
10:00 161.1 F 9.7
12:00 184.0 F 13.3
14:00 205.7 F 4.8
16:00 218.5 F 8.3
18:00 228.6 F 12.3
20:00 230.4 F 8.2
22:00 241.2 F 13.1
24:00 240.6 F 11.2
26:00 252.3 F 17.1
28:00 262.8 F 23.3
30:00 272.7 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.1
1:00 271.9 F 29.6
2:00 272.7 F 29.6
3:00 271.4 F 29.1
4:00 272.6 F 30.0

Min 271.3 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.7 F 0.0
5:00 218.1 F 0.0
10:00 210.2 F 0.0
15:00 192.2 F 0.0
20:00 187.5 F 0.0
25:00 188.5 F 0.0
30:00 193.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:16
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:08:40
CYCLE COMPLETE

02/20/2025 11:38 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____