
Midmark M11 - v1.0.5
Total Cycles: 3266

Name: RITTER2
Sterilizer ID: V2313228
12/29/2023 11:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.9 F 0.0
2:00 148.1 F 0.9
4:00 160.3 F 4.1
6:00 188.0 F 9.4
8:00 212.5 F 4.9
10:00 232.3 F 10.9
12:00 235.3 F 9.4
14:00 236.4 F 8.7
16:00 251.7 F 16.6
18:00 265.3 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 271.8 F 28.8
2:00 272.5 F 29.2
3:00 272.9 F 29.5
4:00 272.7 F 29.8

Min 271.6 F 28.8
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 229.7 F 0.0
Dry Cycle Aborted

FILLING: 1:20
HEATING: 19:32
STERILIZING: 4:00
VENTING: 2:48
DRYING: 2:52
TOTAL CYCLE: 00:30:31

12/29/2023 11:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____