

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000477  
15:01 12/02/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:11  
STERILIZING 4:28  
136.5 C 226kPa 4:28  
136.4 C 233kPa 4:58  
136.4 C 233kPa 5:28  
136.3 C 233kPa 5:58  
136.3 C 232kPa 6:28  
136.4 C 233kPa 6:58  
136.3 C 233kPa 7:28  
136.4 C 232kPa 7:58  
136.4 C 232kPa 7:58  
VENTING 7:58  
AIR DRYING 8:54  
CYCLE COMPLETE 23:54  
-----

Digital Signature #  
7D1895ABA79FAEC5  
-----