

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000115  
14:10 27/02/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:09  
PRESSURIZING 3:11  
STERILIZING 4:14  
136.0 C 220kPa 4:14  
136.2 C 223kPa 4:44  
136.3 C 224kPa 5:14  
136.3 C 224kPa 5:44  
136.4 C 225kPa 6:14  
136.3 C 224kPa 6:44  
136.4 C 221kPa 7:14  
136.4 C 223kPa 7:44  
136.4 C 223kPa 7:44  
VENTING 7:44  
AIR DRYING 8:37  
CYCLE COMPLETE 9:47  
-----

Digital Signature #  
AF4D237A5722282E  
-----