
Midmark M11 - v1.0.0
Total Cycles: 3229

Name: RITTER 1
Sterilizer ID: V2178198
05/15/2023 08:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.6 F 0.0
2:00 174.5 F 0.0
4:00 173.6 F 2.0
6:00 215.1 F 6.5
8:00 215.2 F 11.8
10:00 226.9 F 6.6
12:00 238.2 F 11.9
14:00 236.2 F 9.0
16:00 240.4 F 9.9
18:00 246.6 F 13.5
20:00 258.9 F 20.1
22:00 270.0 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.7 F 29.0
2:00 271.8 F 28.3
3:00 272.1 F 28.9
4:00 272.7 F 28.9

Min 271.4 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 233.0 F 0.0
10:00 221.3 F 0.0
15:00 223.8 F 0.0
20:00 225.8 F 0.0
25:00 231.1 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 22:56
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:01:01
CYCLE COMPLETE

05/15/2023 09:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____