
Midmark M11 - v1.0.0
Total Cycles: 2952

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2023 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.4 F 0.0
2:00 189.6 F 2.2
4:00 194.7 F 6.5
6:00 227.5 F 12.2
8:00 228.9 F 7.7
10:00 240.9 F 13.2
12:00 239.3 F 10.9
14:00 237.3 F 9.3
16:00 251.7 F 16.3
18:00 263.8 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.4 F 29.0
2:00 271.6 F 28.8
3:00 272.4 F 29.0
4:00 271.6 F 28.7

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 231.9 F 0.0
10:00 227.2 F 0.0
15:00 229.0 F 0.0
20:00 228.2 F 0.0
25:00 223.4 F 0.0
30:00 222.3 F 0.0

Drying Complete

FILLING: 1:39
HEATING: 19:58
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:50
CYCLE COMPLETE

02/28/2023 10:49 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____