
Midmark M11 - v1.0.0
Total Cycles: 1527

Name: RITTER 1
Sterilizer ID: V2178198
01/13/2022 04:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.2 F 0.0
2:00 193.0 F 0.2
4:00 213.0 F 1.4
6:00 224.8 F 4.5
8:00 235.5 F 8.5
10:00 246.0 F 13.4
12:00 242.9 F 11.5
14:00 240.4 F 10.3
16:00 238.6 F 9.5
18:00 254.1 F 17.1
20:00 266.9 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 272.5 F 28.4
2:00 272.1 F 28.4
3:00 272.4 F 28.2
4:00 272.1 F 28.4

Min 271.6 F 27.8
Max 272.9 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.0 F 0.0
5:00 234.6 F 0.0
10:00 240.2 F 0.0
Dry Cycle Aborted

FILLING: 1:18
HEATING: 21:23
STERILIZING: 4:00
VENTING: 3:03
DRYING: 10:40
TOTAL CYCLE: 00:40:24

01/13/2022 05:03 PM
ACCEPTED: X
REJECTED:
Operator: U02