
Midmark M11 - v1.0.0
Total Cycles: 1096

Name: RITTER 1
Sterilizer ID: V2178198
08/19/2021 08:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.1 F 0.0
2:00 168.8 F 2.3
4:00 183.5 F 6.0
6:00 221.3 F 11.7
8:00 225.9 F 6.4
10:00 237.3 F 11.8
12:00 235.0 F 8.9
14:00 238.9 F 9.7
16:00 247.6 F 14.0
18:00 260.2 F 20.9
20:00 271.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.2 F 28.5
2:00 272.4 F 29.0
3:00 271.6 F 28.2
4:00 272.4 F 29.0

Min 271.2 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 236.0 F 0.0
10:00 198.2 F 0.0
15:00 189.4 F 0.0
20:00 189.5 F 0.0
25:00 203.5 F 0.0
30:00 189.7 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

08/19/2021 09:09 AM
ACCEPTED: X
REJECTED:
Operator: U02