
Midmark M11 - v1.0.5
Total Cycles: 2425

Name: RITTER2
Sterilizer ID: V2313228
05/29/2023 09:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.9 F 0.0
2:00 173.2 F 2.1
4:00 172.5 F 5.7
6:00 188.5 F 10.4
8:00 212.4 F 5.1
10:00 225.0 F 9.8
12:00 233.9 F 8.7
14:00 241.6 F 12.7
16:00 241.3 F 11.4
18:00 254.1 F 17.9
20:00 265.5 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.5 F 29.7
2:00 271.6 F 29.1
3:00 271.2 F 29.4
4:00 272.5 F 29.7

Min 271.0 F 28.9
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 215.8 F 0.0
10:00 230.6 F 0.0
15:00 223.7 F 0.0
20:00 222.6 F 0.0
25:00 219.9 F 0.0
30:00 217.4 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 21:48
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:25
CYCLE COMPLETE

05/29/2023 10:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____