
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000157
13:53 17/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:42
PRESSURIZING 1:45
STERILIZING 2:42
136.0 C 220kPa 2:42
136.1 C 223kPa 3:12
136.3 C 224kPa 3:42
136.2 C 222kPa 4:12
136.4 C 223kPa 4:42
136.4 C 222kPa 5:12
136.4 C 222kPa 5:42
136.3 C 222kPa 6:12
136.3 C 222kPa 6:12
VENTING 6:12
AIR DRYING 7:07
CYCLE COMPLETE 7:14

Digital Signature #
9C590B790DE8DA21
