
Midmark M11 - v1.0.5
Total Cycles: 1644

Name: RITTER2
Sterilizer ID: V2313228
10/24/2022 11:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.9 F 0.0
2:00 176.8 F 3.4
4:00 185.6 F 7.7
6:00 207.3 F 12.7
8:00 219.9 F 7.1
10:00 232.7 F 12.2
12:00 234.9 F 9.5
14:00 237.0 F 9.2
16:00 247.9 F 14.7
18:00 259.8 F 21.4
20:00 270.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.4 F 29.6
2:00 271.4 F 28.9
3:00 272.3 F 29.8
4:00 272.0 F 29.3

Min 271.3 F 28.9
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.9 F 0.0
5:00 212.3 F 0.0
10:00 230.4 F 0.0
15:00 229.4 F 0.0
20:00 227.9 F 0.0
25:00 227.5 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 20:47
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:33
CYCLE COMPLETE

10/24/2022 12:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____