

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4579

Name: RITTER 1  
Sterilizer ID: V2178198  
05/24/2024 09:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.3 F 0.0  
2:00 173.6 F 2.2  
4:00 176.4 F 5.7  
6:00 194.1 F 9.3  
8:00 219.4 F 9.5  
10:00 221.2 F 7.3  
12:00 231.0 F 11.6  
14:00 231.4 F 7.8  
16:00 243.5 F 13.2  
18:00 241.7 F 11.5  
20:00 253.4 F 17.3  
22:00 263.7 F 23.4  
24:00 272.7 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.7 F 29.4  
2:00 272.1 F 29.0  
3:00 271.3 F 28.6  
4:00 272.2 F 29.3

Min 271.3 F 28.5  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.6 F 0.0  
5:00 208.0 F 0.0  
10:00 224.6 F 0.0  
15:00 237.9 F 0.0  
20:00 203.0 F 0.0  
25:00 195.6 F 0.0  
30:00 190.6 F 0.0  
Drying Complete

FILLING: 1:35  
HEATING: 24:23  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:03:40  
CYCLE COMPLETE

05/24/2024 10:27 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_