
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000106
6:57 03/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:57
PRESSURIZING 4:00
STERILIZING 5:43
136.4 C 224kPa 5:43
136.2 C 223kPa 6:13
136.3 C 223kPa 6:43
136.3 C 224kPa 7:13
136.3 C 221kPa 7:43
136.3 C 224kPa 8:13
136.4 C 225kPa 8:43
136.4 C 225kPa 9:13
136.4 C 225kPa 9:13
VENTING 9:13
AIR DRYING 10:13
CYCLE COMPLETE 25:14

Digital Signature #
084AD79D3F29E15A
