
Midmark M11 - v1.0.0
Total Cycles: 1023

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2021 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.9 F 0.0
2:00 194.7 F 5.0
4:00 212.9 F 11.0
6:00 228.8 F 7.3
8:00 240.7 F 11.9
10:00 243.9 F 12.4
12:00 243.2 F 11.7
14:00 257.2 F 19.1
16:00 269.2 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.2 F 28.5
2:00 272.5 F 28.9
3:00 271.4 F 28.1
4:00 272.6 F 28.8

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.3 F 0.0
Dry Cycle Aborted

FILLING: 2:10
HEATING: 16:59
STERILIZING: 4:00
VENTING: 3:12
DRYING: 3:54
TOTAL CYCLE: 00:30:15

07/15/2021 04:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____