

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3147

Name: RITTER2  
Sterilizer ID: V2313228  
11/20/2023 03:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.3 F 0.0  
2:00 164.6 F 2.8  
4:00 191.0 F 7.2  
6:00 209.9 F 11.7  
8:00 221.1 F 6.2  
10:00 232.6 F 11.5  
12:00 234.3 F 9.0  
14:00 239.6 F 10.1  
16:00 247.7 F 14.3  
18:00 259.6 F 21.0  
20:00 270.4 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.7 F 29.6  
2:00 271.5 F 28.9  
3:00 272.0 F 29.7  
4:00 272.6 F 29.4

Min 271.1 F 28.8  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.3 F 0.0  
5:00 224.3 F 0.0  
10:00 216.0 F 0.0  
15:00 215.3 F 0.0  
20:00 209.9 F 0.0  
25:00 210.9 F 0.0  
30:00 212.4 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 20:51  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 00:59:55  
CYCLE COMPLETE

11/20/2023 04:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_