
Midmark M11 - v1.0.5
Total Cycles: 2131

Name: RITTER2
Sterilizer ID: V2313228
03/09/2023 08:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.1 F 0.0
2:00 80.6 F 0.0
4:00 89.3 F 0.0
6:00 128.4 F 0.2
8:00 158.1 F 2.9
10:00 178.8 F 6.4
12:00 197.8 F 10.2
14:00 212.5 F 7.1
16:00 222.5 F 7.2
18:00 235.0 F 11.3
20:00 232.5 F 7.4
22:00 243.6 F 13.1
24:00 242.2 F 11.5
26:00 254.3 F 17.7
28:00 264.9 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.7 F 29.2
2:00 271.4 F 28.8
3:00 272.4 F 29.2
4:00 271.3 F 28.7

Min 271.3 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 202.3 F 0.0
10:00 206.3 F 0.0
15:00 215.8 F 0.0
20:00 218.3 F 0.0
25:00 217.3 F 0.0
30:00 215.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 29:56
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:07:55
CYCLE COMPLETE

03/09/2023 09:33 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____