
Midmark M11 - v1.0.5
Total Cycles: 3373

Name: RITTER2
Sterilizer ID: V2313228
02/09/2024 12:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.4 F 0.0
2:00 150.2 F 1.3
4:00 156.2 F 4.2
6:00 180.7 F 7.9
8:00 200.4 F 12.2
10:00 218.3 F 5.8
12:00 226.0 F 10.4
14:00 230.3 F 8.1
16:00 242.4 F 12.9
18:00 241.7 F 11.9
20:00 254.3 F 18.2
22:00 265.3 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 271.7 F 29.8
2:00 272.6 F 30.0
3:00 271.4 F 29.5
4:00 272.6 F 30.2

Min 270.9 F 28.9
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 218.3 F 0.0
10:00 220.1 F 0.0
15:00 209.3 F 0.0
20:00 206.5 F 0.0
25:00 203.3 F 0.0
30:00 201.9 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:45
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:02:22
CYCLE COMPLETE

02/09/2024 01:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____