

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4415

Name: RITTER2  
Sterilizer ID: V2313228  
03/10/2025 08:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.3 F 0.0  
2:00 70.9 F 0.2  
4:00 96.3 F 0.7  
6:00 110.7 F 3.1  
8:00 142.6 F 6.2  
10:00 168.6 F 10.0  
12:00 189.0 F 7.1  
14:00 203.2 F 5.5  
16:00 215.9 F 8.9  
18:00 228.0 F 12.8  
20:00 229.9 F 8.4  
22:00 241.4 F 13.4  
24:00 240.1 F 11.3  
26:00 252.4 F 17.4  
28:00 263.2 F 23.8  
30:00 272.7 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.4  
1:00 272.8 F 30.0  
2:00 271.8 F 29.5  
3:00 272.7 F 29.9  
4:00 271.8 F 29.5

Min 271.6 F 29.2  
Max 272.9 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.8 F 0.0  
5:00 185.8 F 0.0  
10:00 193.5 F 0.0  
15:00 193.4 F 0.0  
20:00 195.7 F 0.0  
25:00 194.9 F 0.0  
30:00 192.2 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 30:17  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:08:28  
CYCLE COMPLETE

03/10/2025 09:56 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_