
Midmark M11 - v1.0.0
Total Cycles: 4772

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2024 03:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.7 F 0.0
2:00 168.8 F 3.5
4:00 190.6 F 7.9
6:00 209.7 F 12.9
8:00 223.6 F 6.4
10:00 236.6 F 11.2
12:00 232.9 F 7.7
14:00 245.4 F 12.6
16:00 244.4 F 12.2
18:00 256.9 F 18.7
20:00 268.1 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.2 F 28.4
2:00 271.8 F 28.5
3:00 271.8 F 28.2
4:00 271.9 F 28.6

Min 271.2 F 27.8
Max 272.4 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 218.3 F 0.0
10:00 202.8 F 0.0
15:00 214.7 F 0.0
20:00 202.0 F 0.0
25:00 196.9 F 0.0
30:00 195.8 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 21:14
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:00:15
CYCLE COMPLETE

07/26/2024 04:36 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____