

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000696  
9:18 16/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:29  
PRESSURIZING 2:36  
STERILIZING 3:37  
136.0 C 220kPa 3:37  
136.0 C 230kPa 4:07  
136.3 C 231kPa 4:37  
136.4 C 233kPa 5:07  
136.1 C 230kPa 5:37  
136.1 C 231kPa 6:07  
136.4 C 233kPa 6:37  
136.4 C 232kPa 7:07  
136.4 C 232kPa 7:07  
VENTING 7:07  
AIR DRYING 8:01  
CYCLE COMPLETE 8:07  
-----

Digital Signature #  
68CAD04FE4FD80DE  
-----