
Midmark M11 - v1.0.5
Total Cycles: 3822

Name: RITTER2
Sterilizer ID: V2313228
07/29/2024 01:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.4 F 0.0
2:00 97.4 F 0.0
4:00 119.0 F 0.7
6:00 151.0 F 3.4
8:00 177.1 F 6.3
10:00 210.4 F 9.3
12:00 213.3 F 13.3
14:00 225.0 F 6.6
16:00 228.4 F 10.6
18:00 232.2 F 8.2
20:00 240.2 F 11.5
22:00 237.2 F 9.4
24:00 249.2 F 15.1
26:00 259.6 F 21.0
28:00 269.2 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.1 F 29.0
2:00 272.3 F 29.4
3:00 271.3 F 29.1
4:00 271.8 F 29.0

Min 271.2 F 28.7
Max 272.4 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.1 F 0.0
5:00 230.7 F 0.0
10:00 189.5 F 0.0
15:00 185.9 F 0.0
20:00 182.5 F 0.0
25:00 180.8 F 0.0
30:00 178.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:03
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:07:35
CYCLE COMPLETE

07/29/2024 03:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____