
Midmark M11 - v1.0.5
Total Cycles: 1404

Name: RITTER2
Sterilizer ID: V2313228
08/24/2022 07:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.7 F 0.0
2:00 74.9 F 0.0
4:00 104.4 F 0.0
6:00 116.3 F 0.0
8:00 141.9 F 0.7
10:00 183.6 F 2.6
12:00 204.1 F 4.6
14:00 223.3 F 7.4
16:00 230.8 F 11.1
18:00 232.0 F 6.9
20:00 236.0 F 10.7
22:00 237.3 F 8.3
24:00 244.9 F 13.4
26:00 243.1 F 12.1
28:00 254.6 F 18.1
30:00 265.1 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.3 F 29.2
2:00 271.8 F 29.3
3:00 271.9 F 28.9
4:00 272.0 F 29.5

Min 271.5 F 28.6
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 205.9 F 0.0
10:00 216.6 F 0.0
15:00 212.9 F 0.0
20:00 216.6 F 0.0
25:00 217.0 F 0.0
30:00 214.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:57
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:09:57
CYCLE COMPLETE

08/24/2022 08:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____