
Midmark M11 - v1.0.5
Total Cycles: 207

Name: RITTER2
Sterilizer ID: V2313228
07/14/2021 11:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.8 F 0.0
2:00 185.7 F 3.2
4:00 186.4 F 7.4
6:00 204.2 F 12.2
8:00 219.9 F 6.6
10:00 231.6 F 11.5
12:00 234.0 F 8.7
14:00 241.2 F 11.2
16:00 246.2 F 13.7
18:00 258.0 F 20.2
20:00 268.5 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.8 F 29.6
2:00 271.6 F 28.9
3:00 272.6 F 29.8
4:00 272.3 F 29.3

Min 271.5 F 28.9
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 214.3 F 0.0
10:00 213.0 F 0.0
15:00 214.7 F 0.0
20:00 213.5 F 0.0
25:00 212.4 F 0.0
30:00 211.7 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 21:13
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

07/14/2021 12:25 PM
ACCEPTED: X
REJECTED:
Operator: U01