
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005352
11:40 11/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:01
STERILIZING 3:06
136.4 C 220kPa 3:06
136.4 C 225kPa 3:36
136.4 C 224kPa 4:06
136.5 C 225kPa 4:36
136.4 C 222kPa 5:06
136.4 C 224kPa 5:36
136.4 C 225kPa 6:06
136.4 C 222kPa 6:36
136.4 C 222kPa 6:36
VENTING 6:36
AIR DRYING 7:25
CYCLE COMPLETE 7:49

Digital Signature #
96EA212E30F2BFA2
