
Midmark M11 - v1.0.5
Total Cycles: 3855

Name: RITTER2
Sterilizer ID: V2313228
08/14/2024 08:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 70.4 F 0.0
4:00 74.0 F 0.0
6:00 95.2 F 0.7
8:00 131.9 F 3.3
10:00 160.8 F 6.3
12:00 186.4 F 9.6
14:00 204.8 F 13.4
16:00 216.9 F 6.0
18:00 223.9 F 9.7
20:00 231.4 F 11.0
22:00 235.8 F 10.3
24:00 236.0 F 8.8
26:00 246.5 F 14.0
28:00 257.2 F 20.0
30:00 267.5 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.1
1:00 272.7 F 29.7
2:00 271.6 F 29.4
3:00 272.0 F 29.2
4:00 272.5 F 29.8

Min 271.4 F 29.0
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 206.3 F 0.0
10:00 186.7 F 0.0
15:00 182.3 F 0.0
20:00 179.5 F 0.0
25:00 177.9 F 0.0
30:00 176.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:14
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:09:30
CYCLE COMPLETE

08/14/2024 09:20 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____