

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5159

Name: RITTER 1  
Sterilizer ID: V2178198  
12/04/2024 05:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.5 F 0.0  
2:00 163.1 F 0.3  
4:00 180.5 F 1.2  
6:00 206.2 F 5.3  
8:00 222.0 F 10.0  
10:00 228.5 F 7.5  
12:00 238.2 F 11.7  
14:00 236.7 F 9.8  
16:00 236.3 F 8.5  
18:00 250.5 F 15.6  
20:00 262.8 F 22.6  
22:00 272.2 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.6 F 29.1  
2:00 271.8 F 28.4  
3:00 271.6 F 28.7  
4:00 272.5 F 29.1

Min 271.2 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.7 F 0.0  
5:00 208.4 F 0.0  
10:00 230.4 F 0.0  
15:00 230.7 F 0.0  
20:00 223.9 F 0.0  
25:00 218.4 F 0.0  
30:00 204.9 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 22:13  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:00:53  
CYCLE COMPLETE

12/04/2024 06:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_