
Midmark M11 - v1.0.0
Total Cycles: 1556

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2022 03:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.6 F 0.0
2:00 119.0 F 0.0
4:00 128.9 F 2.3
6:00 159.1 F 5.8
8:00 188.3 F 10.3
10:00 209.0 F 4.2
12:00 225.5 F 7.5
14:00 235.0 F 11.9
16:00 232.9 F 8.0
18:00 244.4 F 12.3
20:00 244.0 F 12.1
22:00 256.8 F 18.8
24:00 268.5 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.6 F 28.7
2:00 272.0 F 28.6
3:00 272.0 F 28.2
4:00 272.7 F 28.9

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 228.2 F 0.0
10:00 210.6 F 0.0
15:00 210.4 F 0.0
20:00 219.3 F 0.0
25:00 222.7 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:59
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:03:03
CYCLE COMPLETE

01/24/2022 04:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____