
Midmark M11 - v1.0.0
Total Cycles: 5305

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2025 12:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.1 F 0.0
2:00 153.9 F 2.6
4:00 163.3 F 5.7
6:00 203.5 F 9.7
8:00 220.9 F 6.0
10:00 221.1 F 5.9
12:00 231.9 F 10.0
14:00 237.5 F 9.5
16:00 240.3 F 11.0
18:00 235.8 F 8.6
20:00 249.2 F 14.7
22:00 259.9 F 20.8
24:00 269.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.2
2:00 271.6 F 28.8
3:00 272.5 F 29.3
4:00 271.5 F 28.7

Min 271.5 F 28.7
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.6 F 0.0
5:00 209.3 F 0.0
10:00 226.5 F 0.0
15:00 226.7 F 0.0
20:00 223.9 F 0.0
25:00 214.0 F 0.0
30:00 211.7 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 25:47
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:04:32
CYCLE COMPLETE

01/31/2025 01:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____