
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007595
7:22 26/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:03
STERILIZING 4:05
136.0 C 220kPa 4:05
136.2 C 221kPa 4:35
136.1 C 220kPa 5:05
136.4 C 221kPa 5:35
136.4 C 222kPa 6:05
136.3 C 222kPa 6:35
135.8 C 218kPa 7:05
136.1 C 221kPa 7:35
136.1 C 221kPa 7:35
VENTING 7:35
AIR DRYING 8:24
CYCLE COMPLETE 38:25

Digital Signature #
3CBE9FE2046BC81A
