
Midmark M11 - v1.0.5
Total Cycles: 1023

Name: RITTER2
Sterilizer ID: V2313228
04/25/2022 07:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.2 F 0.0
2:00 73.9 F 0.0
4:00 92.8 F 0.0
6:00 124.4 F 0.0
8:00 154.2 F 1.4
10:00 181.0 F 4.6
12:00 191.7 F 8.4
14:00 207.5 F 12.0
16:00 216.5 F 5.5
18:00 227.0 F 9.6
20:00 236.6 F 10.8
22:00 238.9 F 11.4
24:00 237.5 F 9.7
26:00 250.0 F 15.9
28:00 260.9 F 22.2
30:00 270.9 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 271.9 F 29.1
2:00 272.1 F 29.3
3:00 272.3 F 29.6
4:00 272.4 F 29.9

Min 271.5 F 29.1
Max 272.4 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 200.9 F 0.0
10:00 210.8 F 0.0
15:00 212.9 F 0.0
20:00 214.0 F 0.0
25:00 218.0 F 0.0
30:00 220.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:37
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:08:31
CYCLE COMPLETE

04/25/2022 08:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____