

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006328  
8:49 20/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:37  
PRESSURIZING 2:41  
STERILIZING 3:44  
136.4 C 222kPa 3:44  
136.3 C 222kPa 4:14  
136.4 C 223kPa 4:44  
136.4 C 223kPa 5:14  
136.4 C 222kPa 5:44  
136.1 C 220kPa 6:14  
136.4 C 223kPa 6:44  
136.4 C 223kPa 7:14  
136.4 C 223kPa 7:14  
VENTING 7:14  
AIR DRYING 8:05  
CYCLE COMPLETE 8:26

-----  
Digital Signature #  
1D292C672FE12BAA  
-----