

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007734  
9:15 13/08/2024

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:20  
PRESSURIZING 3:20  
STERILIZING 4:21  
136.4 C 221kPa 4:21  
136.4 C 222kPa 4:51  
136.2 C 224kPa 5:21  
136.4 C 224kPa 5:51  
136.4 C 221kPa 6:21  
136.3 C 222kPa 6:51  
136.4 C 223kPa 7:21  
136.4 C 219kPa 7:51  
136.4 C 219kPa 7:51  
VENTING 7:51  
AIR DRYING 8:45  
CYCLE COMPLETE 9:40

-----  
Digital Signature #  
0BC690878596CFC4  
-----