
Midmark M11 - v1.0.0
Total Cycles: 3571

Name: RITTER 1
Sterilizer ID: V2178198
07/24/2023 03:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.3 F 0.0
2:00 149.7 F 0.1
4:00 170.9 F 2.0
6:00 190.3 F 4.8
8:00 212.7 F 8.6
10:00 219.6 F 12.9
12:00 225.9 F 6.5
14:00 235.8 F 10.8
16:00 236.5 F 7.8
18:00 244.5 F 13.0
20:00 242.7 F 11.6
22:00 254.4 F 17.6
24:00 264.9 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.8
2:00 271.4 F 28.2
3:00 272.5 F 29.0
4:00 271.8 F 28.5

Min 271.3 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 207.1 F 0.0
10:00 228.3 F 0.0
15:00 230.2 F 0.0
20:00 224.7 F 0.0
25:00 223.3 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 25:60
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:04:29
CYCLE COMPLETE

07/24/2023 04:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____