
Midmark M11 - v1.0.5
Total Cycles: 505

Name: RITTER2
Sterilizer ID: V2313228
10/29/2021 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.7 F 0.0
2:00 128.8 F 1.5
4:00 150.7 F 4.4
6:00 178.6 F 8.9
8:00 199.1 F 8.9
10:00 218.4 F 6.9
12:00 231.7 F 12.0
14:00 234.2 F 9.3
16:00 238.4 F 9.8
18:00 247.7 F 14.7
20:00 260.2 F 21.7
22:00 271.4 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.6 F 29.5
2:00 272.4 F 29.7
3:00 272.3 F 29.4
4:00 272.3 F 29.6

Min 271.6 F 29.0
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.6 F 0.0
5:00 230.4 F 0.0
10:00 214.1 F 0.0
15:00 206.5 F 0.0
20:00 214.4 F 0.0
25:00 196.7 F 0.0
30:00 181.4 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 22:36
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:01:06
CYCLE COMPLETE

10/29/2021 02:13 PM
ACCEPTED: X
REJECTED:
Operator: U02