
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000849
13:09 07/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:24
PRESSURIZING 3:28
STERILIZING 4:39
136.0 C 220kPa 4:39
136.3 C 232kPa 5:09
136.1 C 230kPa 5:39
136.4 C 233kPa 6:09
136.4 C 233kPa 6:39
136.1 C 229kPa 7:09
136.1 C 231kPa 7:39
136.4 C 233kPa 8:09
136.4 C 233kPa 8:09
VENTING 8:09
AIR DRYING 9:02
CYCLE COMPLETE 24:02

Digital Signature #
63B2A8465171C626
